UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/599,672	10/05/2006	Ralf Kuehner	E7900.2048/P2048	2692
24998 DICKSTEIN SI	7590 02/02/201 HAPIRO LLP	EXAMINER		
1825 EYE STR	EET NW	TANNER, JOCELIN C		
Washington, D	C 20000-3403		ART UNIT	PAPER NUMBER
			3731	
			MAIL DATE	DELIVERY MODE
			02/02/2010	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

## Advisory Action Before the Filing of an Appeal Brief

Application No.	Applicant(s)		
10/599,672	KUEHNER ET AL.		
Examiner	Art Unit		

	JOCELIN C. TANNER	3731	
The MAILING DATE of this communication appe	ars on the cover sheet with the c	correspondence add	ress
THE REPLY FILED <u>22 January 2010</u> FAILS TO PLACE THIS A	PPLICATION IN CONDITION FOR	R ALLOWANCE.	
1.  The reply was filed after a final rejection, but prior to or on application, applicant must timely file one of the following rapplication in condition for allowance; (2) a Notice of Appe for Continued Examination (RCE) in compliance with 37 C periods:	eplies: (1) an amendment, affidavit al (with appeal fee) in compliance	t, or other evidence, w with 37 CFR 41.31; or	hich places the (3) a Request
a) The period for reply expiresmonths from the mailing b) The period for reply expires on: (1) the mailing date of this Adno event, however, will the statutory period for reply expire la  Examiner Note: If box 1 is checked, check either box (a) or (I)  MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f)	dvisory Action, or (2) the date set forth interthan SIX MONTHS from the mailing on. ONLY CHECK BOX (b) WHEN THE	date of the final rejection	n.
Extensions of time may be obtained under 37 CFR 1.136(a). The date of have been filed is the date for purposes of determining the period of extrumer 37 CFR 1.17(a) is calculated from: (1) the expiration date of the set forth in (b) above, if checked. Any reply received by the Office later may reduce any earned patent term adjustment. See 37 CFR 1.704(b). NOTICE OF APPEAL	on which the petition under 37 CFR 1.13 ension and the corresponding amount of hortened statutory period for reply origin	of the fee. The appropria nally set in the final Offic	ate extension fee e action; or (2) as
<ol> <li>The Notice of Appeal was filed on A brief in compl filing the Notice of Appeal (37 CFR 41.37(a)), or any exter Notice of Appeal has been filed, any reply must be filed with the complex of Appeal has been filed.</li> </ol>	sion thereof (37 CFR 41.37(e)), to	avoid dismissal of the	
AMENDMENTS			
3. The proposed amendment(s) filed after a final rejection, be  (a) They raise new issues that would require further cor  (b) They raise the issue of new matter (see NOTE below	sideration and/or search (see NOT w);	E below);	
<ul><li>(c) They are not deemed to place the application in bett appeal; and/or</li></ul>	er form for appeal by materially rec	lucing or simplifying ti	ne issues for
(d) ☐ They present additional claims without canceling a c	orresponding number of finally reje	cted claims.	
NOTE: <u>See Continuation Sheet</u> . (See 37 CFR 1.11			
4. The amendments are not in compliance with 37 CFR 1.12	* **	mpliant Amendment (I	PTOL-324).
5. Applicant's reply has overcome the following rejection(s):			·
<ol> <li>Newly proposed or amended claim(s) would be allenon-allowable claim(s).</li> </ol>	owable if submitted in a separate, t	•	_
7.  For purposes of appeal, the proposed amendment(s): a) how the new or amended claims would be rejected is prov. The status of the claim(s) is (or will be) as follows: Claim(s) allowed:		be entered and an ex	xplanation of
Claim(s) objected to: Claim(s) rejected: <u>1-24</u> . Claim(s) withdrawn from consideration:			
AFFIDAVIT OR OTHER EVIDENCE			
<ol> <li>The affidavit or other evidence filed after a final action, but because applicant failed to provide a showing of good and was not earlier presented. See 37 CFR 1.116(e).</li> </ol>			
<ol> <li>The affidavit or other evidence filed after the date of filing a entered because the affidavit or other evidence failed to or showing a good and sufficient reasons why it is necessary</li> </ol>	vercome <u>all</u> rejections under appea and was not earlier presented. Se	ıl and/or appellant fails ee 37 CFR 41.33(d)(1)	s to provide a ).
<ol> <li>The affidavit or other evidence is entered. An explanation REQUEST FOR RECONSIDERATION/OTHER</li> </ol>	of the status of the claims after er	ntry is below or attach	ed.
<ol> <li>The request for reconsideration has been considered but <u>See Continuation Sheet.</u></li> </ol>		condition for allowand	ce because:
<ul><li>12. ☐ Note the attached Information <i>Disclosure Statement</i>(s). (</li><li>13. ☐ Other:</li></ul>	PTO/SB/08) Paper No(s)		
/Anhtuan T. Nguyen/	/Jocelin C. Tanner/		
Supervisory Patent Examiner, Art Unit 3731	Examiner, Art Unit 3731		

Continuation of 3. NOTE: The amended claims include limitations not previously considered and that change the scope of the claims, i.e. a change-over magazine that receives the plurality of cylinders and defines the chambers.

Continuation of 11. does NOT place the application in condition for allowance because: Regarding claim 1 (formerly claim 8), the Applicant contends that Reilly fails to disclose a change -over magazine that receives the plurality of supply cylinders, wherein the change-over magazine defines chambers, each of which receives and closely surrounds a respective one of the plurality supply cylinders. The Applicant further contends that the Examiner is unclear and may be relying on the pressurizing unit chambers 20 as disclosing the claimed change-over magazine which could not also receive and closely surround a respective one of themselves. However, Reilly discloses a change-over magazine (10), that can best be seen in Figs. 2A and 2B, wherein the change-over magazine encases the individual supply cylinders (20) and closely surrounds and receives the supply cylinders.